

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
 Data File : VN074699.D
 Acq On : 01 Oct 2022 08:51
 Operator : JC\MD
 Sample : N4882-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-120

Quant Time: Oct 03 02:13:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 30 07:44:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.939	168	499525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	941201	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.621	117	958660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	439026	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	654500	52.280	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.560%
35) Dibromofluoromethane	7.869	113	530619	53.263	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.520%
50) Toluene-d8	10.310	98	2009968	47.966	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.940%
62) 4-Bromofluorobenzene	12.610	95	649323	48.868	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093022\
Data File : VN074699.D
Acq On : 01 Oct 2022 08:51
Operator : JC\MD
Sample : N4882-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-120

Quant Time: Oct 03 02:13:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092922W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 30 07:44:59 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.939 min Scan# 1029
Delta R.T. 0.006 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51




Instrument :
MSVOA_N
ClientSampleId :
MW-120


Tgt Ion:168 Resp: 499525
Ion Ratio Lower Upper
168 100
99 53.2 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 52.280 ug/l
RT: 8.286 min Scan# 1088
Delta R.T. 0.000 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51

Tgt Ion: 65 Resp: 654500
Ion Ratio Lower Upper
65 100
67 53.1 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.828 min Scan# 1180

Delta R.T. -0.000 min

Lab File: VN074699.D

Acq: 01 Oct 2022 08:51

Instrument :

MSVOA_N

ClientSampleId :

MW-120





Tgt Ion:114 Resp: 941201

Ion Ratio Lower Upper

114 100

63 21.9 0.0 40.0

88 14.8 0.0 32.6



#35

Dibromofluoromethane

Concen: 53.263 ug/l

RT: 7.869 min Scan# 1017

Delta R.T. -0.000 min

Lab File: VN074699.D

Acq: 01 Oct 2022 08:51




Tgt Ion:113 Resp: 530619


Ion Ratio Lower Upper

113 100

111 103.9 81.4 122.0

192 17.5 17.0 25.6









#50
Toluene-d8
Concen: 47.966 ug/l
RT: 10.310 min Scan# 1432
Delta R.T. -0.000 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51




Instrument :
MSVOA_N
ClientSampleId :
MW-120


Tgt Ion: 98 Resp: 2009968
Ion Ratio Lower Upper
98 100
100 63.7 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 48.868 ug/l
RT: 12.610 min Scan# 1823
Delta R.T. 0.006 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51

Tgt Ion: 95 Resp: 649323
Ion Ratio Lower Upper
95 100
174 81.5 0.0 167.6
176 78.7 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.621 min Scan# 1
Delta R.T. 0.005 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51

Instrument :
MSVOA_N
ClientSampleId :
MW-120

Tgt Ion:117 Resp: 958660
Ion Ratio Lower Upper
117 100
82 55.1 44.2 66.2
119 30.9 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.551 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN074699.D
Acq: 01 Oct 2022 08:51

Tgt Ion:152 Resp: 439026
Ion Ratio Lower Upper
152 100
115 58.5 28.9 86.7
150 152.8 0.0 356.2

