

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079377.D
 Acq On : 30 Sep 2023 11:22
 Operator : JC\MD
 Sample : VN0930WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930WBL01

Quant Time: Oct 02 01:50:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	138477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	222715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	68663	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115491	61.783	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.560%
35) Dibromofluoromethane	8.171	113	102670	62.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	124.860%#
50) Toluene-d8	10.571	98	260720	43.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.340%
62) 4-Bromofluorobenzene	12.853	95	88783	46.091	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079377.D
Acq On : 30 Sep 2023 11:22
Operator : JC\MD
Sample : VN0930WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01

Quant Time: Oct 02 01:50:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

