

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079387.D
 Acq On : 30 Sep 2023 15:22
 Operator : JC\MD
 Sample : PB155992ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#07

Quant Time: Oct 02 01:56:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	133808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	279821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	282214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	74969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118901	65.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 131.660%#	
35) Dibromofluoromethane	8.171	113	98838	53.101	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.200%	
50) Toluene-d8	10.571	98	373275	54.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.240%	
62) 4-Bromofluorobenzene	12.853	95	103960	47.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.380%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	19161	24.977	ug/l	89
20) Methylene Chloride	5.283	84	10561	5.657	ug/l #	77
43) Isopropyl Acetate	8.694	43	182893	39.170	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079387.D
Acq On : 30 Sep 2023 15:22
Operator : JC\MD
Sample : PB155992ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#07

Quant Time: Oct 02 01:56:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

