

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079388.D
 Acq On : 30 Sep 2023 15:46
 Operator : JC\MD
 Sample : PB155992ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#08

Quant Time: Oct 02 01:56:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	131508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	279865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	279655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118665	66.845	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 133.680%#	
35) Dibromofluoromethane	8.171	113	96895	52.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.100%	
50) Toluene-d8	10.571	98	366786	53.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.320%	
62) 4-Bromofluorobenzene	12.853	95	104586	47.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.940%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	19274	25.564	ug/l	93
20) Methylene Chloride	5.289	84	10972	5.980	ug/l #	90
43) Isopropyl Acetate	8.694	43	215653	46.179	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079388.D
Acq On : 30 Sep 2023 15:46
Operator : JC\MD
Sample : PB155992ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#08

Quant Time: Oct 02 01:56:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

