

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079389.D
 Acq On : 30 Sep 2023 16:10
 Operator : JC\MD
 Sample : PB155992ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#09

Quant Time: Oct 02 01:57:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	127699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	274397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	273172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	72319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117744	68.304	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 136.600%#	
35) Dibromofluoromethane	8.171	113	95169	52.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.280%	
50) Toluene-d8	10.570	98	365921	54.599	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.200%	
62) 4-Bromofluorobenzene	12.853	95	102116	47.767	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.540%	
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	19125	26.123	ug/l	96
20) Methylene Chloride	5.277	84	11639	6.533	ug/l #	82
43) Isopropyl Acetate	8.694	43	203128	44.363	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079389.D
Acq On : 30 Sep 2023 16:10
Operator : JC\MD
Sample : PB155992ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#09

Quant Time: Oct 02 01:57:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

