

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079391.D
 Acq On : 30 Sep 2023 16:58
 Operator : JC\MD
 Sample : PB155992ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#11

Quant Time: Oct 02 01:58:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	135188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	277244	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	280227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	76101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117681	64.486	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 128.980%#		
35) Dibromofluoromethane	8.171	113	97431	52.832	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.660%		
50) Toluene-d8	10.571	98	370326	54.689	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.380%		
62) 4-Bromofluorobenzene	12.853	95	101324	46.910	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 93.820%		
Target Compounds						
						Qvalue
16) Acetone	4.442	43	19188	24.757	ug/l	90
20) Methylene Chloride	5.283	84	11384	6.036	ug/l #	89
43) Isopropyl Acetate	8.694	43	217772	47.073	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079391.D
Acq On : 30 Sep 2023 16:58
Operator : JC\MD
Sample : PB155992ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#11

Quant Time: Oct 02 01:58:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

