

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079392.D
 Acq On : 30 Sep 2023 17:22
 Operator : JC\MD
 Sample : PB155992ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#12

Quant Time: Oct 02 01:58:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	130390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278333	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	276931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	74858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	122209	69.431	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 138.860%#	
35) Dibromofluoromethane	8.171	113	97698	52.769	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.540%	
50) Toluene-d8	10.571	98	364520	53.621	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.240%	
62) 4-Bromofluorobenzene	12.853	95	103395	47.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.360%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	20022	26.784	ug/l	99
20) Methylene Chloride	5.289	84	11931	6.559	ug/l #	80
43) Isopropyl Acetate	8.694	43	269785	58.088	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079392.D
Acq On : 30 Sep 2023 17:22
Operator : JC\MD
Sample : PB155992ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#12

Quant Time: Oct 02 01:58:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

