

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079393.D
 Acq On : 30 Sep 2023 17:46
 Operator : JC\MD
 Sample : PB155992ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#13

Quant Time: Oct 02 01:59:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	274296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	270752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	76599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	115277	67.603	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 135.200%#			
35) Dibromofluoromethane	8.171	113	93044	50.995	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.000%			
50) Toluene-d8	10.571	98	357622	53.380	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.760%			
62) 4-Bromofluorobenzene	12.853	95	97558	45.652	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.300%			
Target Compounds						
16) Acetone	4.436	43	19668	27.158	ug/l	89
18) Methyl Acetate	5.042	43	4584	1.540	ug/l #	72
20) Methylene Chloride	5.283	84	11906	6.756	ug/l #	85
25) 2-Butanone	7.506	43	5370	5.111	ug/l	91
43) Isopropyl Acetate	8.694	43	214939	46.960	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079393.D
Acq On : 30 Sep 2023 17:46
Operator : JC\MD
Sample : PB155992ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#13

Quant Time: Oct 02 01:59:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

