

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079394.D
 Acq On : 30 Sep 2023 18:10
 Operator : JC\MD
 Sample : PB155992ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#14

Quant Time: Oct 02 01:59:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	135974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	287334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	283701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	77967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	119565	65.140	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 130.280%#	
35) Dibromofluoromethane	8.171	113	100576	52.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.240%	
50) Toluene-d8	10.571	98	380858	54.269	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.540%	
62) 4-Bromofluorobenzene	12.853	95	106139	47.413	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	18949	24.307	ug/l #	80
18) Methyl Acetate	5.042	43	4555	1.422	ug/l #	80
20) Methylene Chloride	5.283	84	12206	6.434	ug/l	90
37) Ethyl Acetate	7.577	43	3279	1.116	ug/l #	67
43) Isopropyl Acetate	8.688	43	214333	44.703	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079394.D
Acq On : 30 Sep 2023 18:10
Operator : JC\MD
Sample : PB155992ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#14

Quant Time: Oct 02 01:59:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

