

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079395.D
 Acq On : 30 Sep 2023 18:34
 Operator : JC\MD
 Sample : PB155992ZHE#15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155992ZHE#15

Quant Time: Oct 02 02:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	134101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	279331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	272615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	72275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	119777	66.167	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 132.340%#	
35) Dibromofluoromethane	8.171	113	99616	53.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.220%	
50) Toluene-d8	10.571	98	368293	53.982	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	12.853	95	100425	46.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.300%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	18977	24.683	ug/l	90
20) Methylene Chloride	5.283	84	13803	7.378	ug/l #	78
43) Isopropyl Acetate	8.694	43	214144	45.943	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079395.D
Acq On : 30 Sep 2023 18:34
Operator : JC\MD
Sample : PB155992ZHE#15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155992ZHE#15

Quant Time: Oct 02 02:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

