

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
 Data File : VN079377.D
 Acq On : 30 Sep 2023 11:22
 Operator : JC\MD
 Sample : VN0930WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930WBL01

Quant Time: Oct 02 01:50:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	138477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	222715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	68663	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115491	61.783	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.560%
35) Dibromofluoromethane	8.171	113	102670	62.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	124.860%#
50) Toluene-d8	10.571	98	260720	43.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.340%
62) 4-Bromofluorobenzene	12.853	95	88783	46.091	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093023\
Data File : VN079377.D
Acq On : 30 Sep 2023 11:22
Operator : JC\MD
Sample : VN0930WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01

Quant Time: Oct 02 01:50:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. -0.000 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22




Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01


Tgt Ion:168 Resp: 138477
Ion Ratio Lower Upper
168 100
99 65.1 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 61.783 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. -0.001 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22

Tgt Ion: 65 Resp: 115491
Ion Ratio Lower Upper
65 100
67 51.8 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN079377.D

Acq: 30 Sep 2023 11:22

Instrument :

MSVOA_N

ClientSampleId :

VN0930WBL01





Tgt Ion:114 Resp: 247246

Ion Ratio Lower Upper

114 100

63 21.8 0.0 36.4

88 16.0 0.0 30.6



#35

Dibromofluoromethane

Concen: 62.427 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN079377.D

Acq: 30 Sep 2023 11:22




Tgt Ion:113 Resp: 102670


Ion Ratio Lower Upper

113 100

111 99.6 82.6 123.8

192 17.5 12.7 19.1









#50
Toluene-d8
Concen: 43.174 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22




Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01


Tgt Ion: 98 Resp: 260720
Ion Ratio Lower Upper
98 100
100 64.4 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 46.091 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22

Tgt Ion: 95 Resp: 88783
Ion Ratio Lower Upper
95 100
174 76.2 0.0 143.8
176 73.9 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. -0.001 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01

Tgt Ion:117 Resp: 222715
Ion Ratio Lower Upper
117 100
82 55.8 44.0 66.0
119 32.4 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.800 min Scan# 2014
Delta R.T. 0.006 min
Lab File: VN079377.D
Acq: 30 Sep 2023 11:22

Tgt Ion:152 Resp: 68663
Ion Ratio Lower Upper
152 100
115 61.5 32.3 96.8
150 159.4 0.0 355.8

