

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 01 06:49:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	2443	1.026	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.060%#
35) Dibromofluoromethane	8.165	113	2114	1.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.220%#
50) Toluene-d8	10.565	98	7435	1.056	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.120%#
62) 4-Bromofluorobenzene	12.853	95	2451	0.956	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	1.920%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	1928	1.015	ug/l	83
3) Chloromethane	2.365	50	2217	1.022	ug/l	99
4) Vinyl Chloride	2.512	62	1981	0.943	ug/l	89
5) Bromomethane	2.965	94	1910	1.378	ug/l	93
6) Chloroethane	3.124	64	2029	1.349	ug/l	97
7) Trichlorofluoromethane	3.500	101	3140	0.960	ug/l	94
8) Diethyl Ether	3.965	74	1167	0.962	ug/l	88
9) 1,1,2-Trichlorotrifluo...	4.371	101	1934	1.032	ug/l #	63
10) Methyl Iodide	4.583	142	2446	1.013	ug/l #	80
11) Tert butyl alcohol	5.524	59	1919	4.888	ug/l #	73
12) 1,1-Dichloroethene	4.353	96	1584	0.876	ug/l #	50
13) Acrolein	4.183	56	2550	5.658	ug/l	87
14) Allyl chloride	5.018	41	2333	0.743	ug/l	89
15) Acrylonitrile	5.736	53	4920	4.963	ug/l #	81
16) Acetone	4.436	43	4799	4.692	ug/l	90
17) Carbon Disulfide	4.718	76	6680	1.167	ug/l	100
18) Methyl Acetate	5.030	43	2249	1.009	ug/l #	64
19) Methyl tert-butyl Ether	5.806	73	5319	0.872	ug/l #	83
20) Methylene Chloride	5.277	84	2203	1.060	ug/l #	83
21) trans-1,2-Dichloroethene	5.783	96	1753	0.925	ug/l	86
22) Diisopropyl ether	6.671	45	5639	0.868	ug/l #	96
23) Vinyl Acetate	6.606	43	22441	4.634	ug/l #	92
24) 1,1-Dichloroethane	6.571	63	3358	0.925	ug/l #	83
25) 2-Butanone	7.488	43	6488	4.658	ug/l	91
26) 2,2-Dichloropropane	7.494	77	2902	0.886	ug/l	96
27) cis-1,2-Dichloroethene	7.483	96	2259	0.987	ug/l	92
28) Bromochloromethane	7.824	49	1718	1.060	ug/l #	87
29) Tetrahydrofuran	7.841	42	4006	4.660	ug/l	92
30) Chloroform	7.971	83	3793	1.005	ug/l	99
31) Cyclohexane	8.241	56	6798	1.928	ug/l #	59
32) 1,1,1-Trichloroethane	8.171	97	3120	0.921	ug/l #	53
36) 1,1-Dichloropropene	8.365	75	2420	0.871	ug/l	98
37) Ethyl Acetate	7.559	43	2596	0.911	ug/l #	78
38) Carbon Tetrachloride	8.359	117	2795	0.917	ug/l	90
39) Methylcyclohexane	9.600	83	2486	0.804	ug/l #	88
40) Benzene	8.606	78	8245	0.952	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 01 06:49:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	1490	0.963	ug/l #	66
42) 1,2-Dichloroethane	8.671	62	2669	0.909	ug/l	88
43) Isopropyl Acetate	8.688	43	7949	1.424	ug/l #	80
44) Trichloroethene	9.347	130	1885	0.932	ug/l	87
45) 1,2-Dichloropropane	9.624	63	1676	0.818	ug/l #	44
46) Dibromomethane	9.706	93	1144	0.828	ug/l	87
47) Bromodichloromethane	9.894	83	2754	0.897	ug/l #	93
48) Methyl methacrylate	9.676	41	1824	0.821	ug/l	93
49) 1,4-Dioxane	9.694	88	800	19.552	ug/l #	78
51) 4-Methyl-2-Pentanone	10.447	43	11218	4.134	ug/l	97
52) Toluene	10.629	92	4875	0.923	ug/l	91
53) t-1,3-Dichloropropene	10.829	75	2493	0.797	ug/l	98
54) cis-1,3-Dichloropropene	10.306	75	2720	0.808	ug/l	95
55) 1,1,2-Trichloroethane	11.018	97	1669	0.882	ug/l	90
56) Ethyl methacrylate	10.876	69	2402	0.771	ug/l #	77
57) 1,3-Dichloropropane	11.159	76	3092	0.914	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	6179	4.278	ug/l	98
59) 2-Hexanone	11.194	43	8282	4.084	ug/l	96
60) Dibromochloromethane	11.353	129	1916	0.857	ug/l	91
61) 1,2-Dibromoethane	11.465	107	1795	0.913	ug/l	99
64) Tetrachloroethene	11.106	164	1690	0.993	ug/l #	87
65) Chlorobenzene	11.888	112	5025	0.927	ug/l #	86
66) 1,1,1,2-Tetrachloroethane	11.959	131	1632	0.874	ug/l #	66
67) Ethyl Benzene	11.965	91	8582	0.896	ug/l #	88
68) m/p-Xylenes	12.070	106	5869	1.656	ug/l	95
69) o-Xylene	12.400	106	2398	0.715	ug/l	83
70) Styrene	12.412	104	4489	0.790	ug/l	94
71) Bromoform	12.576	173	1168	0.850	ug/l #	94
73) Isopropylbenzene	12.688	105	6991	0.899	ug/l	97
74) N-amyl acetate	12.494	43	3228	0.916	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.941	83	2233	0.954	ug/l #	95
76) 1,2,3-Trichloropropane	12.994	75	3395	1.671	ug/l	53
77) Bromobenzene	12.976	156	1818	0.969	ug/l	98
78) n-propylbenzene	13.035	91	7045	0.781	ug/l	95
79) 2-Chlorotoluene	13.129	91	5383	0.936	ug/l	92
80) 1,3,5-Trimethylbenzene	13.170	105	4830	0.761	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.735	75	660	0.760	ug/l #	94
82) 4-Chlorotoluene	13.223	91	5174	0.893	ug/l	100
83) tert-Butylbenzene	13.435	119	4909	0.870	ug/l	88
84) 1,2,4-Trimethylbenzene	13.482	105	4760	0.744	ug/l	100
85) sec-Butylbenzene	13.612	105	5795	0.770	ug/l	99
86) p-Isopropyltoluene	13.723	119	4495	0.722	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	3424	0.972	ug/l	96
88) 1,4-Dichlorobenzene	13.729	146	3424	0.963	ug/l	96
89) n-Butylbenzene	14.053	91	4866	0.862	ug/l	97
90) Hexachloroethane	14.335	117	1151	0.911	ug/l	82
91) 1,2-Dichlorobenzene	14.100	146	3610	1.046	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.711	75	423	0.876	ug/l	65
93) 1,2,4-Trichlorobenzene	15.400	180	1340	0.788	ug/l	89
94) Hexachlorobutadiene	15.494	225	861	1.015	ug/l	90
95) Naphthalene	15.641	128	5031	0.892	ug/l	97
96) 1,2,3-Trichlorobenzene	15.847	180	1531	0.899	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084218.D
Acq On : 30 Sep 2024 14:48
Operator : JC\MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Oct 01 06:49:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Oct 01 06:49:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:168 Resp: 160573
Ion Ratio Lower Upper
168 100
99 68.1 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 1.015 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 85 Resp: 1928
Ion Ratio Lower Upper
85 100
87 25.3 17.6 52.9









#3
Chloromethane
Concen: 1.022 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 2217
Ion Ratio Lower Upper
50 100
52 34.2 26.8 40.2



#4
Vinyl Chloride
Concen: 0.943 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 62 Resp: 1981
Ion Ratio Lower Upper
62 100
64 39.0 26.2 39.4









#5
Bromomethane
Concen: 1.378 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 1910
Ion Ratio Lower Upper
94 100
96 84.7 72.9 109.3



#6
Chloroethane
Concen: 1.349 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 64 Resp: 2029
Ion Ratio Lower Upper
64 100
66 31.0 23.5 35.3









#7
Trichlorofluoromethane
Concen: 0.960 ug/l
RT: 3.500 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 3140
Ion Ratio Lower Upper
101 100
103 60.0 51.8 77.6



#8
Diethyl Ether
Concen: 0.962 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 74 Resp: 1167
Ion Ratio Lower Upper
74 100
45 83.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.032 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 1934
Ion Ratio Lower Upper
101 100
85 25.4 36.2 54.4#
151 38.8 59.3 88.9#



#10
Methyl Iodide
Concen: 1.013 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48


Tgt Ion:142 Resp: 2446
Ion Ratio Lower Upper
142 100
127 33.8 35.3 52.9#
141 0.0 11.0 16.6#








#11
Tert butyl alcohol
Concen: 4.888 ug/l
RT: 5.524 min Scan# 607
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 59 Resp: 1919
Ion Ratio Lower Upper
59 100
57 0.0 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 0.876 ug/l
RT: 4.353 min Scan# 408
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 96 Resp: 1584
Ion Ratio Lower Upper
96 100
61 88.5 128.6 192.8#
98 34.3 51.7 77.5#





#13

Acrolein

Concen: 5.658 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 56 Resp: 2550

Ion Ratio Lower Upper

56 100

55 60.3 57.2 85.8



#14

Allyl chloride

Concen: 0.743 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 41 Resp: 2333


Ion Ratio Lower Upper

41 100

39 83.1 59.0 88.6


76 29.9 29.8 44.6








#15
Acrylonitrile
Concen: 4.963 ug/l
RT: 5.736 min Scan# 64
Delta R.T. 0.018 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53 Resp: 4920
Ion Ratio Lower Upper
53 100
52 82.8 66.2 99.4
51 0.0 28.8 43.2#



#16
Acetone
Concen: 4.692 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48


Tgt Ion: 43 Resp: 4799
Ion Ratio Lower Upper
43 100
58 24.9 24.4 36.6








#17
Carbon Disulfide
Concen: 1.167 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 76 Resp: 6680
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9



#18
Methyl Acetate
Concen: 1.009 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 2249
Ion Ratio Lower Upper
43 100
74 6.4 19.2 28.8#





#19

Methyl tert-butyl Ether

Concen: 0.872 ug/l

RT: 5.806 min Scan# 653

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 5319

Ion Ratio Lower Upper

73 100

57 13.9 17.7 26.5#



#20

Methylene Chloride

Concen: 1.060 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 84 Resp: 2203




Ion Ratio Lower Upper


84 100

49 100.6 107.2 160.8#

51 43.3 31.8 47.8


86 64.5 52.9 79.3








#21
trans-1,2-Dichloroethene
Concen: 0.925 ug/l
RT: 5.783 min Scan# 651
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:	96	Resp:	1753
Ion	Ratio	Lower	Upper
96	100		
61	115.2	108.4	162.6
98	55.5	49.8	74.8



#22
Diisopropyl ether
Concen: 0.868 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:	45	Resp:	5639
Ion	Ratio	Lower	Upper
45	100		
43	52.8	42.3	63.5
87	22.2	21.8	32.6
59	16.4	9.6	14.4





#23

Vinyl Acetate

Concen: 4.634 ug/l

RT: 6.606 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 22441

Ion Ratio Lower Upper

43 100

86 13.5 8.6 12.8#



#24

1,1-Dichloroethane

Concen: 0.925 ug/l

RT: 6.571 min Scan# 785

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 63 Resp: 3358


Ion Ratio Lower Upper

63 100

98 0.0 3.4 10.2#

100 0.0 2.0 5.8#









#25
2-Butanone
Concen: 4.658 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 6488
Ion Ratio Lower Upper
43 100
72 22.1 21.4 32.2



#26
2,2-Dichloropropane
Concen: 0.886 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 77 Resp: 2902
Ion Ratio Lower Upper
77 100
97 19.7 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 0.987 ug/l
 RT: 7.483 min Scan# 941
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Instrument :
 MSVOA_N
 ClientSampleId :

Tgt Ion: 96 Resp: 2259




Ion	Ratio	Lower	Upper
96	100		
61	140.3	0.0	301.8
98	59.5	0.0	128.2




#28
 Bromochloromethane
 Concen: 1.060 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. 0.012 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 49 Resp: 1718

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.6
130	61.3	58.0	87.0





#29

Tetrahydrofuran

Concen: 4.660 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	42	Resp:	4006
Ion	Ratio	Lower	Upper
42	100		
72	40.5	34.6	51.8
71	33.4	32.2	48.4



#30

Chloroform

Concen: 1.005 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion:	83	Resp:	3793
Ion	Ratio	Lower	Upper
83	100		
85	62.8	51.0	76.4








#31
Cyclohexane
Concen: 1.928 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. -0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 6798
Ion Ratio Lower Upper
56 100
69 87.4 23.8 35.6#
84 103.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 0.921 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 97 Resp: 3120
Ion Ratio Lower Upper
97 100
99 0.0 51.2 76.8#
61 46.5 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 1.026 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 2443

Ion Ratio Lower Upper

65 100

67 54.6 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 114 Resp: 290146


Ion Ratio Lower Upper

114 100

63 22.3 0.0 43.8

88 16.9 0.0 31.6





#35

Dibromofluoromethane

Concen: 1.105 ug/l

RT: 8.165 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	113	Resp:	2114
Ion	Ratio	Lower	Upper
113	100		
111	106.9	83.3	124.9
192	7.2	13.5	20.3#



#36

1,1-Dichloropropene

Concen: 0.871 ug/l




RT: 8.365 min Scan# 1090


Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	75	Resp:	2420
Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.9	50.7
77	31.9	24.7	37.1





#37

Ethyl Acetate

Concen: 0.911 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 43	Resp: 2596
Ion Ratio	Lower Upper
43	100
61	4.8 10.7 16.1#
70	2.4 8.8 13.2#



#38

Carbon Tetrachloride

Concen: 0.917 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 117	Resp: 2795
Ion Ratio	Lower Upper
117	100
119	86.1 78.7 118.1
121	31.6 24.4 36.6





#39

Methylcyclohexane

Concen: 0.804 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 2486

Ion Ratio Lower Upper

83 100

55 87.2 62.8 94.2

98 59.0 38.8 58.2#



#40

Benzene

Concen: 0.952 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 78 Resp: 8245

Ion Ratio Lower Upper

78 100


77 21.1 19.1 28.7








#41
Methacrylonitrile
Concen: 0.963 ug/l
RT: 7.782 min Scan# 990
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 41 Resp: 1490
Ion Ratio Lower Upper
41 100
39 21.5 43.3 64.9#
67 56.9 56.1 84.1
52 0.0 24.5 36.7#



#42
1,2-Dichloroethane
Concen: 0.909 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 62 Resp: 2669
Ion Ratio Lower Upper
62 100
98 4.9 0.0 18.2





#43

Isopropyl Acetate

Concen: 1.424 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 43	Resp: 7949
Ion Ratio	Lower Upper
43	100
61	13.7 20.4 30.6#
87	7.0 10.3 15.5#



#44

Trichloroethene

Concen: 0.932 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 130	Resp: 1885
Ion Ratio	Lower Upper
130	100
95	87.9 0.0 203.0





#45

1,2-Dichloropropane

Concen: 0.818 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 1676

Ion Ratio Lower Upper

63 100

65 0.0 24.6 37.0#



#46

Dibromomethane

Concen: 0.828 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 93 Resp: 1144


Ion Ratio Lower Upper

93 100

95 95.1 65.0 97.4

174 100.3 71.8 107.8





#47

Bromodichloromethane

Concen: 0.897 ug/l

RT: 9.894 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 83	Resp: 2754
Ion Ratio	Lower Upper
83	100
85	68.7 51.5 77.3
127	0.0 5.7 8.5#



#48

Methyl methacrylate

Concen: 0.821 ug/l




RT: 9.676 min Scan# 1313


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 41	Resp: 1824
Ion Ratio	Lower Upper
41	100
69	86.5 67.1 100.7
39	66.4 45.8 68.6





#49

1,4-Dioxane

Concen: 19.552 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	88	Resp:	800
Ion	Ratio	Lower	Upper
88	100		
43	0.0	25.3	37.9#
58	75.9	56.4	84.6



#50

Toluene-d8

Concen: 1.056 ug/l




RT: 10.565 min Scan# 1464


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	98	Resp:	7435
Ion	Ratio	Lower	Upper
98	100		
100	66.7	52.7	79.1









#51
4-Methyl-2-Pentanone
Concen: 4.134 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 11218
Ion Ratio Lower Upper
43 100
58 37.4 31.6 47.4



#52
Toluene
Concen: 0.923 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 92 Resp: 4875
Ion Ratio Lower Upper
92 100
91 159.8 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 0.797 ug/l
 RT: 10.829 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 2493
 Ion Ratio Lower Upper
 75 100
 77 32.7 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 0.808 ug/l
 RT: 10.306 min Scan# 1420
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 75 Resp: 2720
 Ion Ratio Lower Upper
 75 100
 77 34.6 24.2 36.2
 39 46.3 38.7 58.1









#55
1,1,2-Trichloroethane
Concen: 0.882 ug/l
RT: 11.018 min Scan# 1540
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 97	Resp: 1669
Ion Ratio	Lower Upper
97 100	
83 75.7	71.0 106.4
85 57.4	41.7 62.5
99 68.1	50.1 75.1



#56
Ethyl methacrylate
Concen: 0.771 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 69	Resp: 2402
Ion Ratio	Lower Upper
69 100	
41 64.2	53.3 79.9
39 0.0	28.8 43.2#





#57

1,3-Dichloropropane

Concen: 0.914 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 76 Resp: 3092

Ion Ratio Lower Upper

76 100

78 32.2 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 4.278 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 63 Resp: 6179

Ion Ratio Lower Upper

63 100

106 28.2 21.8 32.6









#59
2-Hexanone
Concen: 4.084 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 8282
Ion Ratio Lower Upper
43 100
58 51.4 27.3 81.8



#60
Dibromochloromethane
Concen: 0.857 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 129 Resp: 1916
Ion Ratio Lower Upper
129 100
127 69.2 38.3 114.9









#61
1,2-Dibromoethane
Concen: 0.913 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 1795
Ion Ratio Lower Upper
107 100
109 94.9 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 0.956 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 95 Resp: 2451
Ion Ratio Lower Upper
95 100
174 71.7 0.0 145.2
176 69.2 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion:117 Resp: 244376

Ion Ratio Lower Upper

117 100

82 60.8 47.2 70.8

119 34.0 25.4 38.0



#64

Tetrachloroethene

Concen: 0.993 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:164 Resp: 1690




Ion Ratio Lower Upper


164 100

166 92.5 101.0 151.4#

129 103.5 80.8 121.2


131 95.1 78.1 117.1








#65
Chlorobenzene
Concen: 0.927 ug/l
RT: 11.888 min Scan# 1170
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:112 Resp: 5025
Ion Ratio Lower Upper
112 100
114 41.6 27.0 40.6#



#66
1,1,1,2-Tetrachloroethane
Concen: 0.874 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:131 Resp: 1632
Ion Ratio Lower Upper
131 100
133 96.9 47.9 143.7
119 0.0 32.8 98.3#





#67

Ethyl Benzene

Concen: 0.896 ug/l

RT: 11.965 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 8582

Ion Ratio Lower Upper

91 100

106 24.3 24.5 36.7#



#68

m/p-Xylenes

Concen: 1.656 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:106 Resp: 5869

Ion Ratio Lower Upper

106 100

91 201.7 167.1 250.7









#69
o-Xylene
Concen: 0.715 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 2398
Ion Ratio Lower Upper
106 100
91 248.5 110.2 330.6



#70
Styrene
Concen: 0.790 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:104 Resp: 4489
Ion Ratio Lower Upper
104 100
78 47.1 42.7 64.1
103 52.2 43.6 65.4





#71

Bromoform

Concen: 0.850 ug/l

RT: 12.576 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:173	Resp:	1168
Ion Ratio	Lower	Upper
173	100	
175	43.1	23.5 70.6
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion:152	Resp:	101963
Ion Ratio	Lower	Upper
152	100	
115	62.6	31.3 93.9
150	159.0	0.0 349.8








#73
Isopropylbenzene
Concen: 0.899 ug/l
RT: 12.688 min Scan# 1825
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:105 Resp: 6991
Ion Ratio Lower Upper
105 100
120 24.1 12.9 38.7



#74
N-ethyl acetate
Concen: 0.916 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 3228
Ion Ratio Lower Upper
43 100
70 35.2 34.4 51.6
55 21.8 19.6 29.4
61 14.8 19.6 29.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 0.954 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 2233
Ion Ratio Lower Upper
83 100
131 0.0 4.9 14.7#
85 65.3 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 1.671 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 75 Resp: 3395
Ion Ratio Lower Upper
75 100
77 124.4 96.9 290.7









#77
Bromobenzene
Concen: 0.969 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 1818
Ion Ratio Lower Upper
156 100
77 232.3 115.1 345.3
158 92.7 48.9 146.8



#78
n-propylbenzene
Concen: 0.781 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 91 Resp: 7045
Ion Ratio Lower Upper
91 100
120 24.2 10.8 32.4





#79

2-Chlorotoluene

Concen: 0.936 ug/l

RT: 13.129 min Scan# 1899

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 5383

Ion Ratio Lower Upper

91 100

126 27.2 15.7 47.1



#80

1,3,5-Trimethylbenzene

Concen: 0.761 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 105 Resp: 4830

Ion Ratio Lower Upper

105 100


120 50.6 23.7 71.1








#81
trans-1,4-Dichloro-2-butene
Concen: 0.760 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 75 Resp: 660
Ion Ratio Lower Upper
75 100
53 99.1 78.0 117.0
89 55.5 35.5 53.3#



#82
4-Chlorotoluene
Concen: 0.893 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 91 Resp: 5174
Ion Ratio Lower Upper
91 100
126 31.3 15.7 47.1









#83
tert-Butylbenzene
Concen: 0.870 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 4909
Ion Ratio Lower Upper
119 100
91 59.0 34.1 102.3
134 18.3 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 0.744 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:105 Resp: 4760
Ion Ratio Lower Upper
105 100
120 44.7 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 0.770 ug/l
 RT: 13.612 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 5795
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 0.722 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion:119 Resp: 4495
 Ion Ratio Lower Upper
 119 100
 134 24.2 12.6 37.8
 91 23.4 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 0.972 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:146	Resp:	3424
Ion	Ratio	Lower Upper
146	100	
111	45.5	21.8 65.3
148	67.0	31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 0.963 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.082 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48


Tgt Ion:146	Resp:	3424
Ion	Ratio	Lower Upper
146	100	
111	45.5	21.2 63.6
148	67.0	31.9 95.9








#89
n-Butylbenzene
Concen: 0.862 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 91 Resp: 4866
Ion Ratio Lower Upper
91 100
92 52.1 26.6 79.7
134 21.4 12.3 36.9



#90
Hexachloroethane
Concen: 0.911 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 117 Resp: 1151
Ion Ratio Lower Upper
117 100
201 54.3 34.4 103.2





#91

1,2-Dichlorobenzene

Concen: 1.046 ug/l

RT: 14.100 min Scan# 2066

Delta R.T. -0.006 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:146	Resp:	3610
Ion	Ratio	Lower Upper
146	100	
111	39.9	22.0 66.0
148	57.1	32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.876 ug/l




RT: 14.711 min Scan# 2169


Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion: 75	Resp:	423
Ion	Ratio	Lower Upper
75	100	
155	51.3	38.1 114.5
157	59.1	49.6 149.0








#93
1,2,4-Trichlorobenzene
Concen: 0.788 ug/l
RT: 15.400 min Scan# 21
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:180 Resp: 1340
Ion Ratio Lower Upper
180 100
182 103.0 47.0 141.0
145 41.3 16.4 49.4




#94
Hexachlorobutadiene
Concen: 1.015 ug/l
RT: 15.494 min Scan# 2302
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48



Tgt Ion:225 Resp: 861
Ion Ratio Lower Upper
225 100
223 52.7 32.4 97.2
227 60.7 32.4 97.0





#95

Naphthalene

Concen: 0.892 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084218.D





Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:128	Resp:	5031
Ion Ratio	Lower	Upper
128	100	
127	11.5	10.5 15.7
129	11.3	8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 0.899 ug/l

RT: 15.847 min Scan# 2362

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:180	Resp:	1531
Ion Ratio	Lower	Upper
180	100	
182	101.0	48.2 144.6
145	37.5	17.3 51.9

