

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084222.D
 Acq On : 30 Sep 2024 16:35
 Operator : JC\MD
 Sample : VN0930WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0930WBL01

Quant Time: Oct 01 07:15:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108875	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	129308	48.235	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.480%
35) Dibromofluoromethane	8.171	113	100036	48.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	10.565	98	373040	49.428	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.860%
62) 4-Bromofluorobenzene	12.847	95	127575	46.400	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084222.D
Acq On : 30 Sep 2024 16:35
Operator : JC\MD
Sample : VN0930WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0930WBL01

Quant Time: Oct 01 07:15:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

