

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084225.D
 Acq On : 30 Sep 2024 18:30
 Operator : JC\MD
 Sample : P4116-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D5-20240918

Quant Time: Oct 01 07:15:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165245	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118066	48.189	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.380%		
35) Dibromofluoromethane	8.165	113	89303	47.842	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.680%		
50) Toluene-d8	10.565	98	339801	49.479	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.960%		
62) 4-Bromofluorobenzene	12.847	95	106796	42.686	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	85.380%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.371	101	12677	6.572	ug/l	97
12) 1,1-Dichloroethene	4.348	96	3117	1.675	ug/l #	90
27) cis-1,2-Dichloroethene	7.488	96	3202	1.359	ug/l	78
30) Chloroform	7.965	83	4050	1.043	ug/l #	82
44) Trichloroethene	9.353	130	208236	105.541	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084225.D
Acq On : 30 Sep 2024 18:30
Operator : JC\MD
Sample : P4116-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D5-20240918

Quant Time: Oct 01 07:15:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

