

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093025\
 Data File : VN087997.D
 Acq On : 30 Sep 2025 12:53
 Operator : JC\MD
 Sample : Q3241-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 92625

Quant Time: Oct 01 04:42:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	156388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	345490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	340247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	156977	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	159407	60.567	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.140%
35) Dibromofluoromethane	8.153	113	134039	53.435	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.860%
50) Toluene-d8	10.547	98	438013	49.654	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.829	95	150303	47.668	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.340%
Target Compounds						
16) Acetone	4.418	43	65293	60.296	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093025\
Data File : VN087997.D
Acq On : 30 Sep 2025 12:53
Operator : JC\MD
Sample : Q3241-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
92625

Quant Time: Oct 01 04:42:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

