

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100118\
 Data File : VN051631.D
 Acq On : 1 Oct 2018 12:25
 Operator : MD\SY
 Sample : J5088-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OF-1

Quant Time: Oct 02 15:00:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Sep 29 01:07:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	99679	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	467266	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	397735	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	213718	30.79	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.63%
60) 4-Bromofluorobenzene	12.40	95	124701	20.39	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	67.97%
63) Toluene-d8	10.09	98	523809	27.91	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	93.03%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100118\
Data File : VN051631.D
Acq On : 1 Oct 2018 12:25
Operator : MD\SY
Sample : J5088-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OF-1

Quant Time: Oct 02 15:00:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Sep 29 01:07:29 2018
Response via : Initial Calibration

