

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100118\
 Data File : VN051633.D
 Acq On : 1 Oct 2018 13:14
 Operator : MD\SY
 Sample : J5090-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

MMDadoda
 10/2/2018 3:03:56 PM

Quant Time: Oct 02 01:32:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Sep 29 01:07:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	91056	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	427821	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	359942	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	192266	30.33	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	101.10%		
60) 4-Bromofluorobenzene	12.40	95	110460	19.96	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	66.53%		
63) Toluene-d8	10.09	98	490283	28.87	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	96.23%		
Target Compounds						
8) Diethyl Ether	3.41	74	1558m	0.463	ug/l	Ovalue
18) Methylene Chloride	4.55	84	7388	1.169	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100118\
Data File : VN051633.D
Acq On : 1 Oct 2018 13:14
Operator : MD\SY
Sample : J5090-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Manual Integrations
APPROVED

MMDadoda
10/2/2018 3:03:56 PM

Quant Time: Oct 02 01:32:17 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Sep 29 01:07:29 2018
Response via : Initial Calibration

