

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100118\
 Data File : VN051634.D
 Acq On : 1 Oct 2018 13:39
 Operator : MD\SY
 Sample : J5090-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Oct 02 01:30:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Sep 29 01:07:29 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	94378	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	450555	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	374189	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	203369	30.95	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.17%
60) 4-Bromofluorobenzene	12.40	95	117772	20.47	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	68.23%
63) Toluene-d8	10.09	98	523702	29.66	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.87%

Target Compounds

					Ovalue	
8) Diethyl Ether	3.42	74	1398	0.401	ug/l	94
18) Methylene Chloride	4.55	84	9274	1.416	ug/l #	86
24) Methyl tert-Butyl Ether	5.05	73	80971	5.406	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100118\
Data File : VN051634.D
Acq On : 1 Oct 2018 13:39
Operator : MD\SY
Sample : J5090-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Oct 02 01:30:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Sep 29 01:07:29 2018
Response via : Initial Calibration

