

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100119\
 Data File : VN058455.D
 Acq On : 1 Oct 2019 11:28
 Operator : JC/SP
 Sample : K5107-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-9-30-19

Quant Time: Oct 02 06:02:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	408779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	669703	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	574614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	220444	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	298425	49.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.92%	
35) Dibromofluoromethane	7.57	113	210977	51.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.80%	
50) Toluene-d8	10.09	98	838395	52.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.28%	
62) 4-Bromofluorobenzene	12.41	95	260788	44.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100119\
Data File : VN058455.D
Acq On : 1 Oct 2019 11:28
Operator : JC/SP
Sample : K5107-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-9-30-19

Quant Time: Oct 02 06:02:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

