

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100119\
 Data File : VN058466.D
 Acq On : 1 Oct 2019 16:00
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Oct 02 07:24:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	416868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	678199	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	570690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	218251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	301334	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
35) Dibromofluoromethane	7.57	113	214081	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
50) Toluene-d8	10.08	98	853232	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	12.41	95	265902	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
Target Compounds						
32) 1,1,1-Trichloroethane	7.55	97	12839	1.780	ug/l #	87
64) Tetrachloroethene	10.63	164	47015	12.967	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100119\
Data File : VN058466.D
Acq On : 1 Oct 2019 16:00
Operator : JC/SP
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Oct 02 07:24:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

