

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100119\
 Data File : VN058468.D
 Acq On : 1 Oct 2019 17:47
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

MMDadoda
 10/3/2019 12:35:55 PM

Quant Time: Oct 02 07:30:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	399098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	652847	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	211699	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	291649	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
35) Dibromofluoromethane	7.57	113	203597	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
50) Toluene-d8	10.08	98	817741	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
62) 4-Bromofluorobenzene	12.41	95	255372	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
32) 1,1,1-Trichloroethane	7.56	97	5603m	0.811	ug/l	
64) Tetrachloroethene	10.63	164	24291	6.896	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100119\
Data File : VN058468.D
Acq On : 1 Oct 2019 17:47
Operator : JC/SP
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

MMDadoda
10/3/2019 12:35:55 PM

Quant Time: Oct 02 07:30:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

