

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100119\
 Data File : VN058469.D
 Acq On : 1 Oct 2019 18:09
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Oct 02 07:29:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	398431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	652018	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	540534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	200294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	291491	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
35) Dibromofluoromethane	7.57	113	204792	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
50) Toluene-d8	10.09	98	799336	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
62) 4-Bromofluorobenzene	12.41	95	246737	43.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.18%	
Target Compounds						
32) 1,1,1-Trichloroethane	7.55	97	4852	0.704	ug/l #	42
64) Tetrachloroethene	10.63	164	19280	5.614	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100119\
Data File : VN058469.D
Acq On : 1 Oct 2019 18:09
Operator : JC/SP
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Oct 02 07:29:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

