

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100120\
 Data File : VN063770.D
 Acq On : 1 Oct 2020 14:34
 Operator : JC/MD
 Sample : L4168-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 02 02:49:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	224858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	407648	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	412969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	171298	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	187484	50.47	ug/l	0.00
Spiked Amount						
			Recovery	=	100.94%	
35) Dibromofluoromethane	7.55	113	135577	47.93	ug/l	0.00
Spiked Amount						
			Recovery	=	95.86%	
50) Toluene-d8	10.06	98	516369	48.16	ug/l	0.00
Spiked Amount						
			Recovery	=	96.32%	
62) 4-Bromofluorobenzene	12.38	95	204452	50.72	ug/l	0.00
Spiked Amount						
			Recovery	=	101.44%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100120\
Data File : VN063770.D
Acq On : 1 Oct 2020 14:34
Operator : JC/MD
Sample : L4168-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 02 02:49:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

