

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100120\
 Data File : VN063780.D
 Acq On : 1 Oct 2020 18:33
 Operator : JC/MD
 Sample : PB132009ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132009ZHE#02

Quant Time: Oct 02 03:13:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	216257	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	404852	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	415870	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	168788	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	184020	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
35) Dibromofluoromethane	7.55	113	131945	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	10.06	98	523978	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	12.38	95	205295	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
Target Compounds						
16) Acetone	3.79	43	22262	20.242	ug/l	91
18) Methyl Acetate	4.30	43	4185	1.466	ug/l #	74
20) Methylene Chloride	4.50	84	6964	2.857	ug/l	88
43) Isopropyl Acetate	8.13	43	78168	11.630	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100120\
Data File : VN063780.D
Acq On : 1 Oct 2020 18:33
Operator : JC/MD
Sample : PB132009ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132009ZHE#02

Quant Time: Oct 02 03:13:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

