

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100120\
 Data File : VN063781.D
 Acq On : 1 Oct 2020 18:57
 Operator : JC/MD
 Sample : PB132009ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132009ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 10/2/2020 3:11:52 PM

Quant Time: Oct 02 03:15:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210825	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	405786	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	420841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	172747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	183729	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
35) Dibromofluoromethane	7.55	113	133777	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	10.06	98	518651	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	12.38	95	209199	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
Target Compounds						
16) Acetone	3.78	43	22578	21.058	ug/l	95
18) Methyl Acetate	4.28	43	4127m	1.483	ug/l	
20) Methylene Chloride	4.51	84	6771	2.850	ug/l	95
43) Isopropyl Acetate	8.13	43	77067	11.439	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100120\
Data File : VN063781.D
Acq On : 1 Oct 2020 18:57
Operator : JC/MD
Sample : PB132009ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132009ZHE#03

Manual Integrations
APPROVED

MMDadoda
10/2/2020 3:11:52 PM

Quant Time: Oct 02 03:15:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

