

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100120\
 Data File : VN063784.D
 Acq On : 1 Oct 2020 20:09
 Operator : JC/MD
 Sample : PB132009ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132009ZHE#06

Quant Time: Oct 02 03:19:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	209702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406177	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	413089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	171188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	181284	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
35) Dibromofluoromethane	7.55	113	131395	46.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.24%	
50) Toluene-d8	10.06	98	509455	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
62) 4-Bromofluorobenzene	12.38	95	200544	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
Target Compounds						
16) Acetone	3.78	43	22345	20.952	ug/l	97
18) Methyl Acetate	4.29	43	4650	1.679	ug/l	99
20) Methylene Chloride	4.51	84	6973	2.947	ug/l	92
43) Isopropyl Acetate	8.13	43	77577	11.504	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100120\
Data File : VN063784.D
Acq On : 1 Oct 2020 20:09
Operator : JC/MD
Sample : PB132009ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132009ZHE#06

Quant Time: Oct 02 03:19:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

