

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100120\
 Data File : VN063804.D
 Acq On : 2 Oct 2020 5:20
 Operator : JC/MD
 Sample : L4236-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Manual Integrations
 APPROVED

MMDadoda
 10/2/2020 3:12:16 PM

Quant Time: Oct 02 05:56:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	392718	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	386050	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	159801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	180381	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
35) Dibromofluoromethane	7.55	113	128752	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	10.06	98	492202	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
62) 4-Bromofluorobenzene	12.38	95	189662	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.16	62	92690	36.529	ug/l	97
6) Chloroethane	2.67	64	41603	27.439	ug/l	97
11) Tert butyl alcohol	4.74	59	9692	26.026	ug/l #	81
12) 1,1-Dichloroethene	3.70	96	332961	161.949	ug/l	99
16) Acetone	3.78	43	3358	3.234	ug/l #	74
20) Methylene Chloride	4.51	84	2439m	1.122	ug/l	
24) 1,1-Dichloroethane	5.80	63	91563	20.378	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	50645	19.122	ug/l	87
39) Methylcyclohexane	9.05	83	2241	0.565	ug/l #	88
40) Benzene	8.00	78	9880	0.869	ug/l	97
42) 1,2-Dichloroethane	8.09	62	39218	8.085	ug/l	99
44) Trichloroethene	8.80	130	2123	0.713	ug/l	90
49) 1,4-Dioxane	9.16	88	194072	4239.087	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100120\
Data File : VN063804.D
Acq On : 2 Oct 2020 5:20
Operator : JC/MD
Sample : L4236-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Manual Integrations
APPROVED

MMDadoda
10/2/2020 3:12:16 PM

Quant Time: Oct 02 05:56:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

