

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068777.D
 Acq On : 1 Oct 2021 12:23
 Operator : JC/MD
 Sample : VN1001MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1001MBL01

Quant Time: Oct 02 03:57:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	517321	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	890549	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	727756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	249656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	360020	49.948	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.900%
35) Dibromofluoromethane	8.024	113	285636	51.622	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.240%
50) Toluene-d8	10.443	98	1061543	52.501	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.000%
62) 4-Bromofluorobenzene	12.734	95	320490	45.617	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068777.D
 Acq On : 1 Oct 2021 12:23
 Operator : JC/MD
 Sample : VN1001MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1001MBL01

Quant Time: Oct 02 03:57:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

