

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068778.D
 Acq On : 1 Oct 2021 12:48
 Operator : JC/MD
 Sample : VN1001WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1001WBL01

Quant Time: Oct 02 03:57:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	504240	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	871005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	685459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	256450	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	355160	50.552	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.100%
35) Dibromofluoromethane	8.024	113	279460	51.639	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.280%
50) Toluene-d8	10.443	98	1017996	51.477	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.960%
62) 4-Bromofluorobenzene	12.733	95	311569	45.342	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068778.D
Acq On : 1 Oct 2021 12:48
Operator : JC/MD
Sample : VN1001WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1001WBL01

Quant Time: Oct 02 03:57:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

