

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068780.D
 Acq On : 1 Oct 2021 13:56
 Operator : JC/MD
 Sample : M3915-18 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW4-20210922

Manual Integrations
 APPROVED

MMDadoda
 10/4/2021 5:51:14 PM

Quant Time: Oct 02 03:58:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	499366	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	863811	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	713538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	253394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	346739	49.835	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.660%
35) Dibromofluoromethane	8.024	113	270616	50.422	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.840%
50) Toluene-d8	10.443	98	1064756	54.290	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.580%
62) 4-Bromofluorobenzene	12.734	95	309243	45.378	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.760%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.221	101	2793m	0.510	ug/l	Qvalue
44) Trichloroethene	9.217	130	423639	64.158	ug/l	84
64) Tetrachloroethene	10.980	164	2450	0.460	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068780.D
Acq On : 1 Oct 2021 13:56
Operator : JC/MD
Sample : M3915-18 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW4-20210922

Manual Integrations
APPROVED

MMDadoda
10/4/2021 5:51:14 PM

Quant Time: Oct 02 03:58:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

