

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068784.D
 Acq On : 1 Oct 2021 15:37
 Operator : JC/MD
 Sample : M3915-04 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-20210921

Quant Time: Oct 02 04:00:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	560153	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	967748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	816598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	288581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	390296	50.008	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.020%		
35) Dibromofluoromethane	8.021	113	300745	50.017	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 100.040%		
50) Toluene-d8	10.443	98	1187733	54.057	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 108.120%		
62) 4-Bromofluorobenzene	12.734	95	356679	46.718	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 93.440%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.213	101	6472	1.054	ug/l	# 40
12) 1,1-Dichloroethene	4.178	96	2048	0.389	ug/l	# 80
27) cis-1,2-Dichloroethene	7.321	96	2277	0.299	ug/l	49
44) Trichloroethene	9.217	130	823985	111.386	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068784.D
Acq On : 1 Oct 2021 15:37
Operator : JC/MD
Sample : M3915-04 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137-20210921

Quant Time: Oct 02 04:00:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

