

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100121\  
 Data File : VN068790.D  
 Acq On : 1 Oct 2021 18:08  
 Operator : JC/MD  
 Sample : M3915-19  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 RE109D1-20210922

Quant Time: Oct 02 04:01:55 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092921W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 29 23:31:35 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.088	168	474473	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	829681	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	742138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	246776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	336868	50.956	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.920%
35) Dibromofluoromethane	8.024	113	261692	50.765	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.520%
50) Toluene-d8	10.443	98	1045848	55.520	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	111.040%
62) 4-Bromofluorobenzene	12.733	95	318029	48.587	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.180%
Target Compounds						
						Qvalue
44) Trichloroethene	9.217	130	73760	11.630	ug/l	88
64) Tetrachloroethene	10.979	164	3307	0.597	ug/l	# 85
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100121\  
 Data File : VN068790.D  
 Acq On : 1 Oct 2021 18:08  
 Operator : JC/MD  
 Sample : M3915-19  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 RE109D1-20210922

Quant Time: Oct 02 04:01:55 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092921W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 29 23:31:35 2021  
 Response via : Initial Calibration

