

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068793.D
 Acq On : 1 Oct 2021 19:24
 Operator : JC/MD
 Sample : M4040-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-RR-03(SA)

Quant Time: Oct 02 04:02:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	478694	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	827680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	685632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	242895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	341416	51.189	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 102.380%		
35) Dibromofluoromethane	8.024	113	268332	52.179	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 104.360%		
50) Toluene-d8	10.443	98	1018175	54.182	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 108.360%		
62) 4-Bromofluorobenzene	12.733	95	300929	46.086	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 92.180%		
Target Compounds						Qvalue
30) Chloroform	7.823	83	5197	0.440	ug/l	93
40) Benzene	8.458	78	7733	0.326	ug/l	99
47) Bromodichloromethane	9.764	83	1977	0.210	ug/l #	93
65) Chlorobenzene	11.776	112	3867	0.247	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068793.D
Acq On : 1 Oct 2021 19:24
Operator : JC/MD
Sample : M4040-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-RR-03(SA)

Quant Time: Oct 02 04:02:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

