

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068794.D
 Acq On : 1 Oct 2021 19:49
 Operator : JC/MD
 Sample : M4040-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-RR-50

Quant Time: Oct 02 04:03:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	461346	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	799551	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	657225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	233069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	326659	50.818	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.640%
35) Dibromofluoromethane	8.024	113	260051	52.347	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.700%
50) Toluene-d8	10.443	98	969028	53.380	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.760%
62) 4-Bromofluorobenzene	12.733	95	290447	46.046	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.100%
Target Compounds						
						Qvalue
40) Benzene	8.466	78	7851	0.343	ug/l	98
65) Chlorobenzene	11.776	112	4042	0.270	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068794.D
Acq On : 1 Oct 2021 19:49
Operator : JC/MD
Sample : M4040-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-RR-50

Quant Time: Oct 02 04:03:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

