

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
 Data File : VN068795.D
 Acq On : 1 Oct 2021 20:14
 Operator : JC/MD
 Sample : M4040-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 02 04:03:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 29 23:31:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	496196	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	857223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	679450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	244031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	347435	50.254	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.500%	
35) Dibromofluoromethane	8.021	113	278007	52.197	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.400%	
50) Toluene-d8	10.443	98	1019004	52.357	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.720%	
62) 4-Bromofluorobenzene	12.734	95	301816	44.629	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 89.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100121\
Data File : VN068795.D
Acq On : 1 Oct 2021 20:14
Operator : JC/MD
Sample : M4040-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 02 04:03:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092921W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 29 23:31:35 2021
Response via : Initial Calibration

