

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084232.D
 Acq On : 01 Oct 2024 12:18
 Operator : JC\MD
 Sample : VN1001MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1001MBL01

Quant Time: Oct 02 01:59:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95011	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116492	51.441	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.880%
35) Dibromofluoromethane	8.165	113	89744	50.591	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.565	98	323058	49.500	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.847	95	104637	44.009	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084232.D
Acq On : 01 Oct 2024 12:18
Operator : JC\MD
Sample : VN1001MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1001MBL01

Quant Time: Oct 02 01:59:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

