

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084236.D
 Acq On : 01 Oct 2024 13:54
 Operator : JC\MD
 Sample : P4116-24DL 2X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D5-20240918DL

Quant Time: Oct 02 02:02:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113462	49.065	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.120%		
35) Dibromofluoromethane	8.165	113	87699	49.532	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.060%		
50) Toluene-d8	10.565	98	321801	49.402	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.800%		
62) 4-Bromofluorobenzene	12.847	95	103054	43.426	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	86.860%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.365	101	5938	3.262	ug/l	90
44) Trichloroethene	9.353	130	95879	51.232	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084236.D
Acq On : 01 Oct 2024 13:54
Operator : JC\MD
Sample : P4116-24DL 2X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D5-20240918DL

Quant Time: Oct 02 02:02:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

