

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084237.D
 Acq On : 01 Oct 2024 14:18
 Operator : JC\MD
 Sample : P4116-27DL 2X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP05-20240918DL

Quant Time: Oct 02 02:02:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	240932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119905	49.070	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.140%		
35) Dibromofluoromethane	8.165	113	92977	49.793	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.580%		
50) Toluene-d8	10.565	98	342052	49.790	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.580%		
62) 4-Bromofluorobenzene	12.847	95	108557	43.375	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	86.760%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	6117	3.180	ug/l #	86
44) Trichloroethene	9.353	130	99484	50.405	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084237.D
Acq On : 01 Oct 2024 14:18
Operator : JC\MD
Sample : P4116-27DL 2X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP05-20240918DL

Quant Time: Oct 02 02:02:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

