

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084239.D
 Acq On : 01 Oct 2024 15:06
 Operator : JC\MD
 Sample : P4116-25DL 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D6-20240918DL

Quant Time: Oct 02 07:05:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275229	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94565	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116703	50.148	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.300%	
35) Dibromofluoromethane	8.171	113	91852	50.622	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.240%	
50) Toluene-d8	10.565	98	332288	49.776	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.560%	
62) 4-Bromofluorobenzene	12.847	95	107886	44.361	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.720%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.365	101	1526	0.833	ug/l #	43
44) Trichloroethene	9.353	130	69928	36.461	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084239.D
Acq On : 01 Oct 2024 15:06
Operator : JC\MD
Sample : P4116-25DL 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D6-20240918DL

Quant Time: Oct 02 07:05:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

