

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084244.D
 Acq On : 01 Oct 2024 17:06
 Operator : JC\MD
 Sample : P4209-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-1

Quant Time: Oct 02 02:05:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	272200	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	234906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115369	50.109	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.220%	
35) Dibromofluoromethane	8.165	113	89442	49.842	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.680%	
50) Toluene-d8	10.565	98	329347	49.884	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.760%	
62) 4-Bromofluorobenzene	12.847	95	101823	42.334	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	84.660%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084244.D
Acq On : 01 Oct 2024 17:06
Operator : JC\MD
Sample : P4209-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-1

Quant Time: Oct 02 02:05:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

