

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084246.D
 Acq On : 01 Oct 2024 17:54
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 02 02:06:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122720	51.130	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.260%		
35) Dibromofluoromethane	8.165	113	93269	50.474	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.940%		
50) Toluene-d8	10.565	98	342691	50.407	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.820%		
62) 4-Bromofluorobenzene	12.847	95	107163	43.268	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 86.540%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084246.D
Acq On : 01 Oct 2024 17:54
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 02 02:06:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

