

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084252.D
 Acq On : 01 Oct 2024 20:18
 Operator : JC\MD
 Sample : P4226-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT172S1-20240924

Quant Time: Oct 02 02:08:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119400	50.959	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.920%			
35) Dibromofluoromethane	8.165	113	91260	49.718	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.440%			
50) Toluene-d8	10.565	98	334780	49.573	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.140%			
62) 4-Bromofluorobenzene	12.847	95	104708	42.560	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 85.120%			
Target Compounds						
44) Trichloroethene	9.347	130	62673	32.302	ug/l	99
64) Tetrachloroethene	11.106	164	2387	1.444	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084252.D
Acq On : 01 Oct 2024 20:18
Operator : JC\MD
Sample : P4226-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT172S1-20240924

Quant Time: Oct 02 02:08:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

