

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084255.D
 Acq On : 01 Oct 2024 21:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 02:11:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.591	0.604	-2.2	91	0.00
3 P	Chloromethane	0.675	0.686	-1.6	94	0.00
4 C	Vinyl Chloride	0.654	0.684	-4.6#	95	0.00
5 T	Bromomethane	0.431	0.417	3.2	89	0.00
6 T	Chloroethane	0.468	0.426	9.0	94	0.00
7 T	Trichlorofluoromethane	1.019	1.069	-4.9	93	0.00
8 T	Diethyl Ether	0.378	0.388	-2.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.584	0.599	-2.6	94	0.00
10 T	Methyl Iodide	0.752	0.794	-5.6	91	0.00
11 T	Tert butyl alcohol	0.122	0.118	3.3	89	0.00
12 CM	1,1-Dichloroethene	0.563	0.588	-4.4#	93	0.00
13 T	Acrolein	0.140	0.134	4.3	87	0.00
14 T	Allyl chloride	0.978	0.934	4.5	87	0.00
15 T	Acrylonitrile	0.309	0.327	-5.8	95	0.00
16 T	Acetone	0.318	0.268	15.7	73	0.00
17 T	Carbon Disulfide	1.782	1.710	4.0	92	0.00
18 T	Methyl Acetate	0.694	0.728	-4.9	95	0.00
19 T	Methyl tert-butyl Ether	1.900	1.997	-5.1	91	0.00
20 T	Methylene Chloride	0.647	0.667	-3.1	94	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.612	-3.7	91	0.00
22 T	Diisopropyl ether	2.022	2.178	-7.7	93	0.00
23 T	Vinyl Acetate	1.500	1.613	-7.5	90	0.00
24 P	1,1-Dichloroethane	1.131	1.210	-7.0	94	0.00
25 T	2-Butanone	0.434	0.426	1.8	87	0.00
26 T	2,2-Dichloropropane	1.020	0.816	20.0	70	0.00
27 T	cis-1,2-Dichloroethene	0.713	0.754	-5.8	94	0.00
28 T	Bromochloromethane	0.505	0.472	6.5	88	0.00
29 T	Tetrahydrofuran	0.268	0.279	-4.1	91	0.00
30 C	Chloroform	1.175	1.226	-4.3#	94	0.00
31 T	Cyclohexane	1.098	1.048	4.6	91	0.00
32 T	1,1,1-Trichloroethane	1.055	1.109	-5.1	93	0.00
33 S	1,2-Dichloroethane-d4	0.741	0.737	0.5	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.330	0.333	-0.9	95	0.00
36 T	1,1-Dichloropropene	0.479	0.500	-4.4	91	0.00
37 T	Ethyl Acetate	0.491	0.514	-4.7	95	0.00
38 T	Carbon Tetrachloride	0.525	0.556	-5.9	94	0.00
39 T	Methylcyclohexane	0.533	0.557	-4.5	88	0.00
40 TM	Benzene	1.492	1.571	-5.3	94	0.00
41 T	Methacrylonitrile	0.267	0.289	-8.2	93	0.00
42 TM	1,2-Dichloroethane	0.506	0.538	-6.3	94	0.00
43 T	Isopropyl Acetate	0.962	0.853	11.3	88	0.00
44 TM	Trichloroethene	0.348	0.363	-4.3	93	0.00
45 C	1,2-Dichloropropane	0.353	0.385	-9.1#	95	0.00
46 T	Dibromomethane	0.238	0.264	-10.9	94	0.00
47 T	Bromodichloromethane	0.529	0.572	-8.1	95	0.00
48 T	Methyl methacrylate	0.383	0.410	-7.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
 Data File : VN084255.D
 Acq On : 01 Oct 2024 21:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 02:11:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	99	0.00
50 S	Toluene-d8	1.213	1.254	-3.4	93	0.00
51 T	4-Methyl-2-Pentanone	0.468	0.513	-9.6	93	0.00
52 CM	Toluene	0.910	0.980	-7.7#	94	0.00
53 T	t-1,3-Dichloropropene	0.539	0.561	-4.1	88	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.605	-4.3	88	0.00
55 T	1,1,2-Trichloroethane	0.326	0.356	-9.2	94	0.00
56 T	Ethyl methacrylate	0.537	0.597	-11.2	92	0.00
57 T	1,3-Dichloropropane	0.583	0.617	-5.8	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.267	-7.2	88	0.00
59 T	2-Hexanone	0.349	0.376	-7.7	90	0.00
60 T	Dibromochloromethane	0.385	0.417	-8.3	92	0.00
61 T	1,2-Dibromoethane	0.339	0.353	-4.1	91	0.00
62 S	4-Bromofluorobenzene	0.442	0.451	-2.0	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.348	0.354	-1.7	95	0.00
65 PM	Chlorobenzene	1.109	1.135	-2.3	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.396	-3.7	94	0.00
67 C	Ethyl Benzene	1.961	2.048	-4.4#	93	0.00
68 T	m/p-Xylenes	0.725	0.772	-6.5	92	0.00
69 T	o-Xylene	0.686	0.744	-8.5	93	0.00
70 T	Styrene	1.162	1.279	-10.1	94	0.00
71 P	Bromoform	0.281	0.303	-7.8	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.815	3.923	-2.8	92	0.00
74 T	N-amyl acetate	1.728	1.754	-1.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.147	1.138	0.8	94	0.00
76 T	1,2,3-Trichloropropane	1.021	1.105	-8.2	96	0.00
77 T	Bromobenzene	0.920	0.924	-0.4	95	0.00
78 T	n-propylbenzene	4.421	4.640	-5.0	92	0.00
79 T	2-Chlorotoluene	2.821	2.828	-0.2	91	0.00
80 T	1,3,5-Trimethylbenzene	3.112	3.327	-6.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.426	0.382	10.3	83	0.00
82 T	4-Chlorotoluene	2.840	2.897	-2.0	93	0.00
83 T	tert-Butylbenzene	2.766	2.865	-3.6	92	0.00
84 T	1,2,4-Trimethylbenzene	3.135	3.344	-6.7	93	0.00
85 T	sec-Butylbenzene	3.691	3.870	-4.8	91	0.00
86 T	p-Isopropyltoluene	3.052	3.279	-7.4	91	0.00
87 T	1,3-Dichlorobenzene	1.727	1.727	0.0	94	0.00
88 T	1,4-Dichlorobenzene	1.744	1.717	1.5	94	0.00
89 T	n-Butylbenzene	2.768	2.776	-0.3	87	0.00
90 T	Hexachloroethane	0.620	0.606	2.3	92	0.00
91 T	1,2-Dichlorobenzene	1.693	1.705	-0.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.231	2.5	95	0.00
93 T	1,2,4-Trichlorobenzene	0.834	0.831	0.4	87	0.00
94 T	Hexachlorobutadiene	0.416	0.354	14.9	87	0.00
95 T	Naphthalene	2.764	2.728	1.3	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100124\
Data File : VN084255.D
Acq On : 01 Oct 2024 21:30
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 02 02:11:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.835	0.813	2.6	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6