

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058483.D
 Acq On : 2 Oct 2019 18:29
 Operator : JC/SP
 Sample : K4663-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-01-093019

Quant Time: Oct 03 05:44:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	378762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	635546	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	547485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	195566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287213	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	7.57	113	200400	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
50) Toluene-d8	10.08	98	797560	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	12.41	95	245713	44.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.06%	
Target Compounds						
16) Acetone	3.80	43	13563	7.188	ug/l	99
20) Methylene Chloride	4.53	84	161281	37.040	ug/l	98
43) Isopropyl Acetate	8.15	43	119461	12.330	ug/l #	92
95) Naphthalene	15.14	128	20787	2.148	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058483.D
Acq On : 2 Oct 2019 18:29
Operator : JC/SP
Sample : K4663-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-01-093019

Quant Time: Oct 03 05:44:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

