

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058484.D
 Acq On : 2 Oct 2019 18:51
 Operator : JC/SP
 Sample : K4660-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Oct 03 05:46:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	378049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	639324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	547771	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	203492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	285193	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	7.58	113	200131	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
50) Toluene-d8	10.09	98	803238	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
62) 4-Bromofluorobenzene	12.41	95	248389	44.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.50%	
Target Compounds						
16) Acetone	3.80	43	12392	6.580	ug/l	99
20) Methylene Chloride	4.53	84	87362	20.101	ug/l	99
43) Isopropyl Acetate	8.16	43	130325	13.372	ug/l #	92
78) n-propylbenzene	12.60	91	18569	1.096	ug/l	98
80) 1,3,5-Trimethylbenzene	12.74	105	33607	2.670	ug/l	100
84) 1,2,4-Trimethylbenzene	13.05	105	31894	2.529	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058484.D
Acq On : 2 Oct 2019 18:51
Operator : JC/SP
Sample : K4660-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Oct 03 05:46:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

