

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058485.D
 Acq On : 2 Oct 2019 19:14
 Operator : JC/SP
 Sample : K5148-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-100119

Quant Time: Oct 03 05:49:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	383450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	650611	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	564536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	212916	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287309	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	7.58	113	201595	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	10.09	98	811786	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	12.41	95	252803	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
16) Acetone	3.80	43	15385	8.054	ug/l #	87
20) Methylene Chloride	4.53	84	325164	73.764	ug/l	99
43) Isopropyl Acetate	8.16	43	137850	13.898	ug/l #	91
95) Naphthalene	15.14	128	22851	2.169	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058485.D
Acq On : 2 Oct 2019 19:14
Operator : JC/SP
Sample : K5148-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-100119

Quant Time: Oct 03 05:49:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

