

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058488.D
 Acq On : 2 Oct 2019 20:20
 Operator : JC/SP
 Sample : K5095-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-1

Quant Time: Oct 03 05:58:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	371961	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	628255	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	542342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	209551	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.01	65	282090	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	
35) Dibromofluoromethane	7.57	113	201437	52.31	ug/l	0.00
Spiked Amount			Recovery	=	104.62%	
50) Toluene-d8	10.09	98	802002	53.17	ug/l	0.00
Spiked Amount			Recovery	=	106.34%	
62) 4-Bromofluorobenzene	12.41	95	257679	47.24	ug/l	0.00
Spiked Amount			Recovery	=	94.48%	

Target Compounds

					Qvalue
16) Acetone	3.80	43	39334	21.228	ug/l 97
20) Methylene Chloride	4.53	84	4877637	1140.673	ug/l 90
40) Benzene	8.03	78	216917	12.562	ug/l 97
43) Isopropyl Acetate	8.16	43	185195	19.336	ug/l # 83
65) Chlorobenzene	11.44	112	1122728	104.144	ug/l 99
88) 1,4-Dichlorobenzene	13.37	146	16444	2.510	ug/l 90
91) 1,2-Dichlorobenzene	13.66	146	6958	1.100	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058488.D
Acq On : 2 Oct 2019 20:20
Operator : JC/SP
Sample : K5095-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-1

Quant Time: Oct 03 05:58:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

