

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058489.D
 Acq On : 2 Oct 2019 21:05
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Oct 03 06:00:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	360389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	607012	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	512802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	192903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	267887	50.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.74%	
35) Dibromofluoromethane	7.58	113	187698	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
50) Toluene-d8	10.09	98	741441	50.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.41	95	232790	44.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.36%	
Target Compounds						
20) Methylene Chloride	4.52	84	4452	1.075	ug/l #	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058489.D
Acq On : 2 Oct 2019 21:05
Operator : JC/SP
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Oct 03 06:00:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

