

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100219\
 Data File : VN058493.D
 Acq On : 2 Oct 2019 22:34
 Operator : JC/SP
 Sample : K5131-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190828-COMPOSITE-D

Quant Time: Oct 03 06:28:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	368318	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	615252	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	529531	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	193665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	274297	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	7.58	113	190214	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	10.08	98	768090	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	12.41	95	238683	44.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.38%	
Target Compounds						
16) Acetone	3.80	43	93664	51.049	ug/l	95
20) Methylene Chloride	4.53	84	1416765	334.598	ug/l	99
25) 2-Butanone	6.82	43	21182	8.408	ug/l	94
43) Isopropyl Acetate	8.15	43	102281	10.905	ug/l	94
65) Chlorobenzene	11.44	112	80281	7.627	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100219\
Data File : VN058493.D
Acq On : 2 Oct 2019 22:34
Operator : JC/SP
Sample : K5131-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190828-COMPOSITE-D

Quant Time: Oct 03 06:28:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

