

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN100219\
 Data File : VN058476.D
 Acq On : 2 Oct 2019 15:54
 Operator : JC/SP
 Sample : VN1002MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1002MBL01

Quant Time: Oct 03 05:28:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	351746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	573312	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	480273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	174730	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	257058	50.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.02%	
35) Dibromofluoromethane	7.57	113	181795	51.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.48%	
50) Toluene-d8	10.08	98	716427	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
62) 4-Bromofluorobenzene	12.41	95	211534	42.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.00%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN100219\
Data File : VN058476.D
Acq On : 2 Oct 2019 15:54
Operator : JC/SP
Sample : VN1002MBL01
Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1002MBL01

Quant Time: Oct 03 05:28:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. -0.00 min

Lab File: VN058476.D

Acq: 2 Oct 2019 15:54

Instrument :

MSVOA_N

ClientSampleId :


VN1002MBL01

Tot Ion:168 Resp: 351746

Ion Ratio Lower Upper




168 100

99 63.0 47.7 71.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#33

1,2-Dichloroethane-d4

Concen: 50.009 ug/l

RT: 8.01 min Scan# 1940

Delta R.T. -0.00 min

Lab File: VN058476.D


Acq: 2 Oct 2019 15:54

Tgt Ion: 65 Resp: 257058

Ion Ratio Lower Upper



65 100


67 51.1 0.0 102.4



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.57 min Scan# 2115
Delta R.T. -0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54

Instrument :
MSVOA_N
ClientSampleId :
VN1002MBL01

Tot Ion:114 Resp: 573312
Ion Ratio Lower Upper
114 100
63 22.8 0.0 45.0
88 16.0 0.0 32.2




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VNO
Ion 87.90 (87.60 to 88.60): VNO






#35
Dibromofluoromethane
Concen: 51.738 ug/l
RT: 7.57 min Scan# 1804
Delta R.T. -0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54

Tgt Ion:113 Resp: 181795
Ion Ratio Lower Upper
113 100
111 103.7 83.4 125.2
192 18.0 14.9 22.3



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.047 ug/l
RT: 10.08 min Scan# 2585
Delta R.T. -0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54




Instrument :
MSVOA_N
ClientSampled :
VN1002MBL01


Tot Ion: 98 Resp: 716427
Ion Ratio Lower Upper
98 100
100 63.1 51.0 76.6



#62
4-Bromofluorobenzene
Concen: 42.501 ug/l
RT: 12.41 min Scan# 3307
Delta R.T. 0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54

Tgt Ion: 95 Resp: 211534
Ion Ratio Lower Upper
95 100
174 75.6 0.0 150.6
176 71.4 0.0 146.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54

Instrument :
MSVOA_N
ClientSampleId :
VN1002MBL01

Tot Ion:117 Resp: 480273
Ion Ratio Lower Upper
117 100
82 58.7 47.4 71.0
119 31.9 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3601
Delta R.T. 0.00 min
Lab File: VN058476.D
Acq: 2 Oct 2019 15:54

Tgt Ion:152 Resp: 174730
Ion Ratio Lower Upper
152 100
115 60.5 30.0 90.0
150 156.9 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

